


## Omnicell refrigerator out-of-range troubleshooting



- Make sure the door is closed properly.
- Make sure there is nothing against the sensor.
- If it continues to be out of range, place a note on the refrigerator stating that it is **OUT OF ORDER**.
- Remove medications if necessary.
- Let the charge nurse know.
- Contact maintenance if necessary.
- Contact our Omnicell TSE Jerry (704-467-2386) if necessary.

